



中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Services Co.,Ltd.(Foshan)

Test Report	Security query website: www.fcl-sz.org.cn Security code: 6732363020
Report Number ZFLJ2613271	Page 1, A total of three pages



Customer Information	Entrusting Party	Yiwu Rishang Bags and Suitcases Co.,Ltd
	Entrusting Party Address	No.289 Huaxian Road, Niansanli Industrial Park , Yiwu City, Zhejiang, P.R.C, 322013
Customer Identification Information	Sample Description	: Disposable face mask
	Model	: KN95
	Sample Quantity	: 35 pcs
Nature of The Test	Commission Test	Sample Acceptance Date: 2020-04-10, Date of Issue: 2020-04-14
Inspect/Determine Basis	- GB 2626-2006 《Respiratory protective equipment self-absorption filter respirator against particulate matter》 .	
Comprehensive Test Conclusion	• The sample is checked and the item is determined to be in conformity with the requirements conform to <GB 2626-2006 .> Corresponding technical requirements	
Inspection results and individual decision	Vide infra	





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Services Co.,Ltd.(Foshan)

Samples presented

(The following are the photos of the submitted samples and/or the pasted samples of the self-submitted samples)



Approve :

张志荣

Prepare document :

沈健容





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Services Co.,Ltd.(Foshan)

Test Report	Security query website: www.fcl-sz.org.cn Security code: 6732363020
Report Number ZFLJ2613271	Page 2, A total of three pages



Test item	technical requirements	inspection result
Information to be provided by the manufacturer(GB 2626-2006)		
Should be provided with minimum sales package	According to standard terms 5.15.1	conformity
There should be instructions in Chinese	According to standard terms 5.15.1	conformity
It should include information that the user must know	According to standard terms 5.15.1	conformity
Warnings should be provided about possible problems in use	According to standard terms 5.15.1	conformity
The message should be clear	According to standard terms 5.15.1	conformity
synthetic judgment		conformity





中国认可
国际互认
检测
TESTING
CNAS L1842



Test Report	Security query website: www.fcl-sz.org.cn Security code: 6732363020
Report Number ZFLJ2613271	Page 3, A total of three pages



Test item (unit of measurement) [sample identification]	test method (parameter description)	Technical requirements (note)	test result	determine
Fillet	6.11 (with the abandoned mask bearing tension of 10N, duration of 10s)	There should be no slippage or fracture	No slippage or fracture occurred	conformity
Expiratory resistance (Pa)	Article 6.6[1] (ventilation capacity: 85 L/min)	≤ 250	25.1	conformity
Filtration efficiency (%)	Clause 6.3[2](gas flow rate 85L/min, aerosol particle NaCl)	(KN95) ≥ 95.0	99.2	conformity
Inspiratory resistance (Pa)	GB 2626-2006 clause 6.5[1] (ventilation capacity 85 L/min)	≤ 350	40.3	conformity
Appearance inspection	Clause 6.1~6.2 of GB 2626-2006	The surface of the sample should not be damaged, deformed or have obvious other defects.	conformity	conformity
		Component materials and structures shall be able to withstand normal operating conditions and possible temperature, humidity and mechanical shocks	conformity	
		After temperature and humidity pretreatment and mechanical strength pretreatment, the parts should not	conformity	





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Services Co.,Ltd.(Foshan)

		fall off, damage and deformation		
Combustible (s)	Clause 6.15 of GB 2626-2006	Afterburn time 5	Afterburn time 0	conformity
—————				
Remark	<p>1. The test items in this report are all carried out under the environmental conditions specified by the corresponding standards (unless indicated).</p> <p>2. The entrusted test results are only responsible for incoming samples, and the duplicate report without the "special seal for inspection and testing" of the institution is invalid.</p> <p>3. Unless required by the customer, the uncertainty of measurement results is not taken into account in the test results and conformance determination of this report.</p> <p>4.[1]= the test method is not within the CNAS approved capacity of the laboratory.[2]= this method is not within the scope of CMA qualification recognition and CNAS accreditation of the laboratory, and has no social proof effect.</p>			

===== End of test report =====

